

Something Different.

WATSON BIO LAB
MADE IN JAPAN SINCE 1989



Made in JAPAN

Speedy

Smooth

Slim

Light
Comfortable to use

NEXTY-S

Silent

7 types in total

NT-S100

10-100 μ L

NT-S200

20-200 μ L

NT-S2

0.2-2 μ L

NT-S10

1-10 μ L

NT-S20

2-20 μ L

NT-S1000

100-1000 μ L

NT-S5000

500-5000 μ L



NEXTY-S

Color-coded

NEXTY's push button has different color by each maximum volume.



Eject button

A design that is unobtrusive during dispensing and easy to press during ejection.




Digital indicator

Scale is easy to look at because display is three digit. Can set 1/1000 volume of maximum.



Lock lever

Push button and turbo dial's rotation can be locked by this lock lever. Can prevent the deviations in volume setting.



Autoclavable bottom parts

The nozzle-cone and eject cone can be taken off and autoclaved.

*The nozzle-cone, O-rings, and seal rings can all be autoclaved, which means there's no need to remove the O-rings and seal rings for the autoclaving process.

Speedy!

Push-Button

A precise volume setting dial for accurate minimum scale unit setting. (1/1000)

Speedy

Triple Speed Turbo Dial

1 rotation of triple turbo dial is equal to 3.5 rotations of the push-button dial.

Quite speedy even when making large volume changes.

| Cat. No. | Item |
|----------|---|
| NT-S2 | NEXTY-S2 Single Channel pipette 0.2~2μL |
| NT-S10 | NEXTY-S10 Single Channel pipette 1~10μL |
| NT-S20 | NEXTY-S20 Single Channel pipette 2~20μL |
| NT-S100 | NEXTY-S100 Single Channel pipette 10~100μL |
| NT-S200 | NEXTY-S200 Single Channel pipette 20~200μL |
| NT-S1000 | NEXTY-S1000 Single Channel pipette 100~1000μL |
| NT-S5000 | NEXTY-S5000 Single Channel pipette 500~5000μL |

Slim! Light! NEXTY-S



Light

Light stroke, Less Fatigue

Light stroke and easy handling reduce repetitive stress on the thumb.



Slim

Slim Body, Easy to Grasp

Thanks to its slim body, those who have small hands can easily grasp it, and it realizes stable dispensing.

| Minimal scale | Volume | Accuracy | Precision | Compatible tip | Packaging |
|---------------|---------|----------|-----------|-----------------------------|-----------|
| 0.001 µL | 0.2 µL | ± 15.0% | ≤ 8.0% | 204/207 | 1 unit |
| | 2 µL | ± 2.5% | ≤ 1.0% | | |
| 0.01 µL | 1 µL | ± 5.0% | ≤ 2.5% | 204/ 207 | 1 unit |
| | 10 µL | ± 1.0% | ≤ 0.5% | | |
| 0.01 µL | 2 µL | ± 6.0% | ≤ 3.0% | 705/703/503 | 1 unit |
| | 20 µL | ± 1.0% | ≤ 0.5% | | |
| 0.1 µL | 10 µL | ± 2.0% | ≤ 1.0% | 705/703/503/505/801 | 1 unit |
| | 100 µL | ± 0.8% | ≤ 0.3% | | |
| 0.1 µL | 20 µL | ± 2.0% | ≤ 1.0% | 705/703/503/505/801 | 1 unit |
| | 200 µL | ± 0.8% | ≤ 0.3% | | |
| 1 µL | 100 µL | ± 2.0% | ≤ 0.6% | 502/ 706/804/806/805 | 1 unit |
| | 1000 µL | ± 0.8% | ≤ 0.3% | | |
| 5 µL | 500 µL | ± 2.0% | ≤ 0.6% | 401 | 1 unit |
| | 5000 µL | ± 0.8% | ≤ 0.3% | | |

Note: Each pipette is calibrated with the tip denoted in bold.
Using other compatible tips may result in slight deviations in performance.